RAAD VOOR ACCREDITATIE **R** 

Dutch Accreditation Council RvA PO Box 2768 NL-3500 GT Utrecht

The Dutch Accreditation Council RvA, by law appointed as the national accreditation body for The Netherlands, hereby declares that accreditation has been granted to:

## TRESCAL Zoetermeer B.V. Technical Operations Zoetermeer

The organisation has demonstrated to be able to generate technical valid results in a competent way and work according to a management system.

This accreditation is based on an assessment against the requirements as laid down in EN ISO/IEC 17025:2017.

The accreditation covers the activities as specified in the authorized annex bearing the registration number.

The accreditation is valid provided that the organisation continues to meet the requirements.

The accreditation with registration number:

## K 052

is granted on 12 September 1989

This declaration is valid until **1 March 2026** 

The board of the Dutch Accreditation Council, on its behalf,

r. J.A.W.M. de Haas

Annex to declaration of accreditation (scope of accreditation) Normative document: EN ISO/IEC 17025:2017 Registration number: **K 052** 

## of TRESCAL Zoetermeer B.V. Technical Operations

This annex is valid from: **17-02-2022** to **01-03-2026** 

Replaces annex dated: zie T06

HCS code	Measured quantity, Range	Frequency	CMC <sup>1</sup>	Remarks	Location
TF 0 0	Time and Frequency				
TF 2 1	Frequency				ZTM
	100 kHz		1.10 <sup>-11</sup> . <i>f</i>	Measurement measuring time τ ≥1000 s	
	1 MHz		1·10 <sup>-11</sup> · <i>f</i>		
	5 MHz		1.10 <sup>-11</sup> . <i>f</i>		
	10 MHz		1.10 <sup>-11</sup> . <i>f</i>		
	0.1 Hz – 1 Hz		12 µHz	Measurement. Generation	
	1 Hz – 10 Hz		12 µHz	measuring time ⊤≥20 s	
	10 Hz – 100 Hz		12 µHz – 1.2 µHz		
	100 Hz – 1 kHz		1.2 µHz		
	1 kHz – 10 kHz		1.2 µHz		
	10 kHz – 100 kHz		1.2 µHz		
	100 kHz – 1 MHz		1.2 μHz – 12 μHz		
	1 MHz – 10 MHz		12 µHz – 0.12 mHz		
	10 MHz – 100 MHz		0.12 mHz – 1.2 mHz		
	100 MHz – 1 GHz		1.2 mHz – 12 mHz		
	1 GHz – 3 GHz		12 mHz – 14 mHz		
	3 GHz – 27.5 GHz		1.2 Hz		
TF 2 2	Time interval			Measurement	ZTM
	100 ps – 1 ns		1.2·10 <sup>-9</sup> · <i>T</i>		
	1 ns – 10 ns		1.2·10 <sup>-9</sup> · <i>T</i>		
	10 ns – 100 ns		1.2·10 <sup>-9</sup> · <i>T</i>		
	100 ns – 1µs		1.2·10 <sup>-9</sup> · <i>T</i>		
	1 µs – 10 µs		1.2·10 <sup>-9</sup> · <i>T</i>		
	10 µs – 100 µs		1.2·10 <sup>-9</sup> · <i>T</i>		
	100 µs – 1 ms		1.2·10 <sup>-9</sup> · <i>T</i>		
	1 ms – 10 ms		1.2·10 <sup>-9</sup> · <i>T</i>		

Dutch Accreditation Council RvA

Annex to declaration of accreditation (scope of accreditation) Normative document: EN ISO/IEC 17025:2017 Registration number: **K 052** 

## of TRESCAL Zoetermeer B.V. Technical Operations

This annex is valid from: 17-02-2022 to 01-03-2026

Replaces annex dated: zie T06

HCS code	Measured quantity, Range	Frequency	CMC <sup>1</sup>	Remarks	Location
	10 ms – 100 ms		1.2·10 <sup>-8</sup> · <i>T</i> – 1.2·10 <sup>-6</sup> · <i>T</i>		
	100 ms – 1 s		1.2·10 <sup>-6</sup> · <i>T</i> – 1.2·10 <sup>-5</sup> · <i>T</i>		
	1 s – 10 s		1.2·10 <sup>-5</sup> · <i>T</i> – 1.2·10 <sup>-4</sup> · <i>T</i>		
TF 2 2	Time interval			Measurement	
	0.1 µs – 100 ms		1⋅10 <sup>-6</sup> ⋅ <i>T</i> + 10 ns	Equipment with separated electrical start and stop inputs.	
	100 ms – 1 s		1·10 <sup>-5</sup> · <i>T</i> + 10 ns		
	1 s – 10 s		1⋅10 <sup>-4</sup> ⋅ <i>T</i> + 10 ns		

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>2</sup>	Remarks	Location		
OQ 0 0	Optical quantities						
OQ 1 3	Optical system properties				ZTM		
	Optical wavelength	1511 – 1542 nm	0.2 pm	Generation of wavelength with a wavelength reference cell, fixed wavelengths			
		840 – 860 nm	0.4 pm	Generation of wavelength in			
		1270 – 1650 nm	0.4 pm	reference wavelength meter			
		840 – 860 nm	0.4 pm	Measurement of wavelengt			
		1270 – 1650 nm	0.4 pm	with a reference wavelength meter			
		600 – 1530 nm	300 pm	Measurement of wavelength with an optical spectrum analyser			
		1530 – 1570 nm	50 pm				
		1570 – 1750 nm	300 pm				

<sup>&</sup>lt;sup>2</sup> Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "*Evaluation* of the Uncertainty of Measurement in Calibration".