



Organisme belge d'Accréditation
Belgische Accreditatieinstelling
Belgische Akkreditierungsstelle
Belgian Accreditation Body

EA MLA Signatory

Bijlage bij accreditatie-certificaat
Annexe au certificat d'accréditation
Annex to the accreditation certificate
Beilage zur Akkreditierungszertifikat

001-CAL

EN ISO/IEC 17025:2017

Versie / Version / Version / Fassung	22
Geldigheidsperiode / Validité / Validity / Gültigkeitsdauer	2023-12-08 - 2026-06-01

Maureen Logghe

Voorzitster van het Accreditatiebureau
La Présidente du Bureau d'Accréditation
Chair of the Accreditation Board
Vorsitzende des Akkreditierungsbüro

De accreditatie werd uitgereikt aan / L'accréditation est délivrée à /
The accreditation is granted to / Die akkreditierung wurde erteilt für:

TRESCAL nv
Vosstraat, 200
2600 Antwerpen

Activiteitencentra / Sites d'activités / Sites of activities / Standorte mit aktivitäten:

Locatie 1 - WOMMELGEM	Nijverheidsstraat, 70 2160 Wommelgem
Locatie 2 - WELLIN	Rue Jean Meunier, 2 6922 Wellin
Locatie 3 - LOUVAIN-LA-NEUVE	Rue du Bosquet, 7 1348 Ottignies-Louvain-la-Neuve

Accréditation
Service public fédéral Economie
P.M.E., Classes moyennes et Energie
Bd du Roi Albert II 16 - 1000 Bruxelles
Numéro d'entreprise : 0314.595.348

Accreditatie
Federale Overheidsdienst Economie
K.M.O., Middenstand en Energie
Koning Albert II-laan 16 - 1000 Brussel
Ondernemingsnummer : 0314.595.348

+32 2 277 54 34
belac@economie.fgov.be
www.belac.be

.be

BELAC

BELAC

BELAC

Mass Wommelgem
Calibration and Measurement Capabilities

Mass standards

Measured quantity, instrument or gauge	Range	expanded uncertainty (*)	Remarks	Calibration procedure
Weights and masses	1 mg	0,001 0 mg	For example mass pieces up to grade E1 according to OIML R111-1	KI/02/KC/W.02 KI/02/KC/W.03 KI/02/KC/W.04 KI/02/KC/W.05
	2 mg	0,001 0 mg		
	5 mg	0,001 0 mg		
	10 mg	0,001 0 mg		
	20 mg	0,001 0 mg		
	50 mg	0,001 2 mg		
	100 mg	0,001 6 mg		
	200 mg	0,002 0 mg		
	500 mg	0,002 5 mg		
	1 g	0,003 0 mg		
	2 g	0,004 0 mg		
	5 g	0,005 0 mg		
	10 g	0,007 0 mg		
	20 g	0,008 0 mg		
	50 g	0,010 mg		
	100 g	0,017 mg		
	200 g	0,033 mg		
	500 g	0,080 mg		
	1 kg	0,16 mg		
	2 kg	0,33 mg		
	5 kg	0,80 mg		
10 kg	1,7 mg			
20 kg	10 mg	For example mass pieces up to grade E2 according to OIML R111-1	KI/02/KC/W.02 KI/02/KC/W.03 KI/02/KC/W.04 KI/02/KC/W.05	
50 kg	600 mg	For example mass pieces up to grade M1 according to OIML R111-1	KI/02/KC/W.05	
100 kg	1 000 mg			
150 kg	1 600 mg			

Weighing instruments

Measured quantity, instrument or gauge	Range	expanded uncertainty (*)	Remarks	Calibration procedure
Non automatic weighing machines			Available weights: grade E1: 1 mg to 1 kg grade E2: 1 mg to 10 kg grade F1: 1 g to 20 kg grade M1: 1 g to 500 kg	P2-02-W.001
	1 mg to 1 kg	$5,0 \times 10^{-6} \times m$ <i>minimal 0,002 mg</i>		
	> 1 kg to 10 kg	$2,0 \times 10^{-6} \times m$		
	> 10 kg to 240 kg	$3,0 \times 10^{-6} \times m$		
	> 240 kg to 10 000 kg	$1,1 \times 10^{-4} \times m$		

(*) the smallest uncertainty of measurement the laboratory can provide to its customers, expressed as the expanded uncertainty having a coverage probability of approximately 95%.

Mass Wellin

Calibration and Measurement Capabilities

Mass standards

Measured quantity, instrument	Range	expanded measurement uncertainty (*)	Remarks	Calibration procedure
Weights and masses	1 g	0,3 mg	OIML R111 defined Class M1, M2 and M3 weights or equivalent, using traceable calibrated F1 reference standards while applying the direct comparison method such as defined and elaborated in OIML R111-1 Annex C.	P1-02-W.010
	2 g	0,4 mg		
	5 g	0,5 mg		
	10 g	0,6 mg		
	20 g	0,8 mg		
	50 g	1,0 mg		
	100 g	1,6 mg		
	200 g	3,0 mg		
	500 g	8 mg		
	1 kg	16 mg		
	2 kg	30 mg		
	5 kg	80 mg		
	10 kg	0,16 g		
	20 kg	0,3 g		P1-02-W.008
	50 kg	0,8 g		

(*) the smallest uncertainty of measurement the laboratory can provide to its customers, expressed as the expanded uncertainty having a coverage probability of approximately 95%.